



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/646,899	08/21/2003	John L. Sommer	P0011139.00	1340
27581	7590	03/15/2010		
MEDTRONIC, INC. 710 MEDTRONIC PARKWAY NE MINNEAPOLIS, MN 55432-9924			EXAMINER ALTER, ALYSSA MARGO	
			ART UNIT 3762	PAPER NUMBER
			MAIL DATE 03/15/2010	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/646,899

Applicant(s)

SOMMER ET AL.

Examiner

Alyssa M. Alter

Art Unit

3762

Period for Reply -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☐ Claim(s) ____ is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☐ Claim(s) ____ is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on ____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. ____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date ____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date ____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: ____

DETAILED ACTION

Response to Arguments

Applicant's arguments filed August 31, 2009 have been fully considered but they are not persuasive. The Applicant argues that Ries et al. does not disclose lead electrode array selectivity since Ries et al. does not disclose "an array of lead connector pads". However, Ries et al. does disclose an array of lead connector pads as depicted in figures 11B-11D. Figures 11B-11D depict a lead with multiple lead connector pads and thus an array of lead connector pads. Furthermore, when Ries et al. inserts the lead into the adaptor body there is registration with the teeth or key with a selected lead connector pad. Therefore, the teeth or key, selectively engage with the lead connector pad. As such, the claims remain rejected under Ries et al. Note, the claims do not state that when the lead is rotated that each individual key will come in contact with a different lead contact pad and that the contacts are spaced in the same plane circumferentially around the outside of the lead.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

1. Claims 1-6, 8-10 and 12-13 stand rejected under 35 U.S.C. 102(e) as being anticipated by Ries et al. (US 6,755,694 B2). Ries et al. include an implantable medical device with an upsizing lead adaptor that conforms to industry standards and has an inner lumen and a conductive ring (212 in figure 5) with teeth (214 in figure 5). The examiner considers the teeth to be the resilient key. Therefore, the key is dimensioned to be in registration with a selected one the lead connector pads at the time of insertion into the lumen of the adaptor body. Figures 11B-11D depict a lead with multiple lead connector pads and thus an array of lead connector pads. Furthermore, Ries et al. does disclose lead electrode array selectivity since placement of the teeth or key to engage with one of the a lead connector pad thus provide selectivity. In addition, the adaptor includes interior sealing rings are depicted in figure 5 as 224 and 226.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of

the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alyssa M. Alter whose telephone number is (571)272-4939. The examiner can normally be reached on M-F 8am to 4pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Carl Layno can be reached on (571) 272-4949. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/George R Evanisko/
Primary Examiner, Art Unit 3762

/Alyssa M Alter/
Examiner
Art Unit 3762